

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT
(IECEE) CB SCHEME

CB TEST CERTIFICATE

Product	Battery Charger
Name and address of the applicant	Victron Energy B.V. De Paal 35, 1351 JG Almere The Netherlands
Name and address of the manufacturer	Victron Energy B.V. De Paal 35, 1351 JG Almere The Netherlands
Name and address of the factory	<input checked="" type="checkbox"/> Additional information on page 2
Note: When more than one factory, please report on page 2	
Ratings and principal characteristics	IMPULSE-II 24/6 Smart, IMPULSE-II 24/6 Smart IP65: Input: 220-240V; 1,7A; 50/60Hz; 210W; Output: 24Vdc; 6A IMPULSE-II 24/8 Smart, IMPULSE-II 24/8 Smart IP65 Input: 220-240V; 2,2A; 50/60Hz; 290W; Output: 24Vdc; 8A IMPULSE-II 24/11 Smart Input: 220-240V; 2,7A; 50/60Hz; 380W; Output: 24Vdc; 11A IMPULSE-II 24/12 Smart IP65 Input: 220-240V; 2,9A; 50/60Hz; 415W; Output: 24Vdc; 12A
Trademark (if any)	Victron Energy
Customer's Testing Facility (CTF) Stage used	
Model / Type Ref.	IMPULSE-II 24/6 Smart IMPULSE-II 24/6 Smart IP65 IMPULSE-II 24/8 Smart IMPULSE-II 24/8 Smart IP65 IMPULSE-II 24/11 Smart IMPULSE-II 24/12 Smart IP65
Additional information (if necessary may also be reported on page 2)	<input type="checkbox"/> Additional information on page 2 No evaluation or certification of Electromagnetic compatibility and Risk Management performed.
A sample of the product was tested and found to be in conformity with	IEC 60601-1-11:2015, IEC 60601-1-11:2015/AMD1:2020, IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012, IEC 60601-1:2005/AMD2:2020
As shown in the Test Report Ref. No. which forms part of this Certificate	17PP350-12_0, 17PP350-13_0

This CB Test Certificate is issued by the National Certification Body

Kiwa Nederland B.V.
Wilmersdorf 50, NL-7327 Apeldoorn, AC, Netherlands

Raphael Rader

- Digitally signed | see <http://ca.kiwa-deutschland.de> for more details -